

JCSS

20260619評基第009号
2026年6月24日

登 録 証

富士フイルム和光純薬株式会社 殿

計量法第143条第1項の規定に基づく校正事業者として登録します。

登 録 番 号 0016

事業所の名称 富士フイルム和光純薬株式会社 東京工場

所 在 地 埼玉県川越市大字的場1633番地

登録に係る区分 濃度（詳細は別紙のとおり）

登録の有効期限 2027年10月30日

2026年6月24日

独立行政法人製品評価技術基盤機構

理事長 長谷川 史彦



Effective Date of Accreditation: 2023-10-21

< Calibration Laboratory >

General Field of Calibration: ConcentrationDate of Initial Accreditation of the Field: 2005-12-26Laboratory's permanent facility/On-site Calibration: Laboratory's permanent facilityCalibration and Measurement Capabilities

Calibration Procedures# and Type of Instruments/Materials to be calibrated	Range	Expanded Uncertainty* (Level of Confidence Approximately 95 %)
pH Standard Solutions	Oxalate pH standard solution	1.679
	Phthalate pH standard solution	4.008
	Phosphate Equimolar pH standard solution	6.865
	Phosphate pH standard solution	7.413
	Tetraborate pH standard solution	9.18
	Carbonate pH standard solution	10.012
Standard Solutions except pH Standard Solutions	Aluminum standard solution	1000 mg/L
	Arsenic standard solution	100 mg/L
		1000 mg/L
	Bismuth standard solution	1000 mg/L
	Calcium standard solution	1000 mg/L
	Cadmium standard solution	100 mg/L
		1000 mg/L
	Cobalt standard solution	1000 mg/L
	Chromium standard solution	100 mg/L
		1000 mg/L
	Copper standard solution	100 mg/L
		1000 mg/L
	Iron standard solution	100 mg/L
		1000 mg/L
	Mercury standard solution	100 mg/L
		1000 mg/L
	Potassium standard solution	1000 mg/L
	Magnesium standard solution	1000 mg/L
	Manganese standard solution	100 mg/L
		1000 mg/L
	Sodium standard solution	1000 mg/L
	Nickel standard solution	1000 mg/L
	Lead standard solution	100 mg/L
		1000 mg/L
	Antimony standard solution	1000 mg/L
	Zinc standard solution	100 mg/L
		1000 mg/L
	Barium standard solution	1000 mg/L
	Lithium standard solution	1000 mg/L
	Molybdenum standard solution	1000 mg/L
	Rubidium standard solution	1000 mg/L
	Selenium standard solution	1000 mg/L
	Tin standard solution	1000 mg/L
	Strontium standard solution	1000 mg/L
	Thallium standard	1000 mg/L
	Boron standard solution	1000 mg/L

Effective Date of Accreditation: 2023-10-21

Standard Solutions except pH Standard Solutions	Cesium standard solution	1000 mg/L	0.7 %
	Gallium standard solution	1000 mg/L	0.5 %
	Indium standard solution	1000 mg/L	0.7 %
	Tellurium standard solution	1000 mg/L	1.5 %
	Vanadium standard solution	1000 mg/L	0.7 %
	Silver standard solution	1000 mg/L	0.4 %
	Beryllium standard solution	1000 mg/L	0.7 %
	Titanium standard solution	1000 mg/L	0.4 %
	Silicon standard solution	1000 mg/L	0.8 %
	Yttrium standard solution	1000 mg/L	0.5 %
	Ammonium ion standard solution	1000 mg/L	0.8 %
	Bromide ion standard solution	1000 mg/L	0.6 %
	Chloride ion standard solution	1000 mg/L	0.6 %
	Fluoride ion standard solution	1000 mg/L	0.7 %
	Nitrate ion standard solution	1000 mg/L	0.8 %
		4430 mg/L	0.9 %
	Nitrite ion standard solution	1000 mg/L	0.8 %
		3280 mg/L	0.9 %
	Phosphate ion standard solution	1000 mg/L	0.7 %
		3070 mg/L	0.7 %
	Sulfate ion standard solution	1000 mg/L	0.9 %
		3000 mg/L	1.1 %
	Chlorate ion standard solution	1000 mg/L	0.7 %
	Bromate ion standard solution	2000 mg/L	0.5 %
	Chlorite ion standard solution	1000 mg/L	1.9 %
	Formaldehyde standard solution	1000 mg/L	4.8 %
	Total organic carbon standard solution	1000 mg/L	0.6 %
	Heptaoxyethylene dodecyl ether standard solution	100 mg/L	3.9 %
Standard Solutions except pH Standard Solutions	5 Anion mixture standard solution		
	Fluoride ion standard solution	100 mg/L	0.5 %
	Chloride ion standard solution	200 mg/L	1.3 %
	Nitrite ion standard solution	33 mg/L	1.1 %
	Chlorate ion standard solution	100 mg/L	0.8 %
	Nitrate ion standard solution	443 mg/L	1.3 %
	7 Anion mixture standard solution		
	Fluoride ion standard solution	20 mg/L	2.2 %
	Chloride ion standard solution	20 mg/L	1.7 %
	Nitrite ion standard solution	100 mg/L	1.1 %
	Bromide ion standard solution	100 mg/L	2.3 %
	Nitrate ion standard solution	100 mg/L	1.5 %
	Phosphate ion standard solution	200 mg/L	2.1 %
	Sulfate ion standard solution	100 mg/L	1.7 %

Effective Date of Accreditation: 2023-10-21

Standard Solutions except pH Standard Solutions	25 VOC mixture standard solution		
	** Any combination of constituents listed as below can be supplied.		
	1, 1-Dichloroethylene	1000 mg/L	1.6 %
	Dichloromethane	1000 mg/L	1.2 %
	<i>trans</i> -1, 2- Dichloroethylene	1000 mg/L	1.1 %
	<i>tert</i> -Butyl methyl ether	1000 mg/L	0.9 %
	<i>cis</i> -1, 2-Dichloroethylene	1000 mg/L	1.1 %
	Chloroform	1000 mg/L	0.9 %
	1, 1, 1-Trichloroethane	1000 mg/L	1.1 %
	Carbon tetrachloride	1000 mg/L	1.4 %
	Benzene	1000 mg/L	1.2 %
	1, 2-Dichloroethane	1000 mg/L	1.3 %
	Trichloroethylene	1000 mg/L	1.1 %
	1, 2-Dichloropropane	1000 mg/L	1.0 %
	Bromodichloromethane	1000 mg/L	1.0 %
	<i>cis</i> -1, 3- Dichloropropene	1000 mg/L	2.8 %
	Toluene	1000 mg/L	1.0 %
	<i>trans</i> -1, 3-Dichloropropene	1000 mg/L	4.1 %
	Tetrachloroethylene	1000 mg/L	1.0 %
	Dibromochloromethane	1000 mg/L	1.1 %
	<i>o</i> -Xylene	1000 mg/L	1.1 %
	Tribromomethane	1000 mg/L	1.1 %
	1, 4-Dichlorobenzene	1000 mg/L	1.0 %
	1, 4-dioxane	1000 mg/L	2.5 %
	<i>p</i> -Xylene	1000 mg/L	0.7 %
	<i>m</i> -Xylene	1000 mg/L	0.8 %
	1, 1, 2-Trichloroethane	1000 mg/L	0.8 %
	2 Musty odor substances mixture standard solution		
	2-Methylisoborneol	100 mg/L	2.2 %
	Geosmin	100 mg/L	1.9 %
	6 Phenols mixture standard solution		
	2-Chlorophenol	1000 mg/L	2.1 %
	Phenol	1000 mg/L	2.2 %
	2, 6-Dichlorophenol	1000 mg/L	2.3 %
	2, 4-Dichlorophenol	1000 mg/L	2.3 %
	2, 4, 6-Trichlorophenol	1000 mg/L	2.4 %
	4-Chlorophenol	1000 mg/L	2.3 %
	4 Haloacetic acid mixture standard solution		
	Chloroacetic acid	1000 mg/L	2.8%
	Dichloroacetic acid	1000 mg/L	3.0 %
	Bromoacetic acid	1000 mg/L	2.8 %
	Trichloroacetic acid	1000 mg/L	3.1 %
	5 Anionic surfactants mixture standard solution		
	Sodium Decylbenzenesulfonate	100 mg/L	4.8 %
	Sodium Undecylbenzenesulfonate	100 mg/L	4.9 %
	Sodium Dodecylbenzenesulfonate	100 mg/L	4.9 %
	Sodium Tridecylbenzenesulfonate	100 mg/L	5.2 %
	Sodium Tetradecylbenzenesulfonate	100 mg/L	5.0 %

#All Calibration Procedures are in-house procedures developed by this laboratory.

* relative value

Effective Date of Accreditation: 2023-10-21

<Reference Material Producer>

Category: Concentration

Date of Initial Accreditation of the Field: 2007-10-24

Type: Certified Reference Material

The Approach Used to Assign a Property Value: Measurement by a Single Method in a Single Laboratory (ISO 17034:2016 7.12.3 NOTE 1 d))

Sub-category and Property		Range	Expanded Uncertainty ^{*1} (Level of Confidence Approximately 95 %)	Characterization Techniques ^{*2}
pH Standard Solutions	Oxalate pH standard solution	1.679	0.008	Glass electrode method
	Phthalate pH standard solution	4.008	0.009	
	Phosphate equimolal pH standard solution	6.865	0.010	
	Phosphate pH standard solution	7.413	0.009	
	Tetraborate pH standard solution	9.18	0.012	
	Carbonate pH standard solution	10.012	0.014	
Standard Solutions except pH Standard Solutions	Aluminum standard solution	1000 mg/L	0.5 %	Titration
	Arsenic standard solution	1000 mg/L	0.5 %	Titration
		100 mg/L	1.3 %	
	Bismuth standard solution	1000 mg/L	1.0 %	Titration
	Calcium standard solution	1000 mg/L	0.6 %	Titration
	Cadmium standard solution	100 mg/L	0.9 %	Titration
		1000 mg/L	0.5 %	
	Cobalt standard solution	1000 mg/L	0.6 %	Titration
	Chromium standard solution	100 mg/L	0.8 %	Titration
		1000 mg/L	0.7 %	
	Copper standard solution	100 mg/L	1.0 %	Titration
		1000 mg/L	0.7 %	
	Iron standard solution	100 mg/L	1.1 %	Titration
		1000 mg/L	0.7 %	
	Mercury standard solution	100 mg/L	0.8 %	Titration
		1000 mg/L	0.6 %	
	Potassium standard solution	1000 mg/L	0.7 %	IC
	Magnesium standard solution	1000 mg/L	0.6 %	Titration
	Manganese standard solution	100 mg/L	0.9 %	Titration
		1000 mg/L	0.4 %	
	Sodium standard solution	1000 mg/L	0.7 %	IC
	Nickel standard solution	1000 mg/L	0.5 %	Titration
	Lead standard solution	100 mg/L	1.3 %	Titration
		1000 mg/L	0.7 %	
	Antimony standard solution	1000 mg/L	0.4 %	Titration
	Zinc standard solution	100 mg/L	0.9 %	Titration
		1000 mg/L	0.5 %	
	Barium standard solution	1000 mg/L	0.7 %	Titration
	Lithium standard solution	1000 mg/L	0.6 %	IC
	Molybdenum standard solution	1000 mg/L	0.6 %	Titration
	Rubidium standard solution	1000 mg/L	0.9 %	IC
	Selenium standard solution	1000 mg/L	1.0 %	Titration
	Tin standard solution	1000 mg/L	0.7 %	Titration
	Strontium standard solution	1000 mg/L	0.5 %	Titration
	Thallium standard	1000 mg/L	0.6 %	Titration
	Boron standard solution	1000 mg/L	0.5 %	Titration
	Cesium standard solution	1000 mg/L	0.7 %	IC
	Gallium standard solution	1000 mg/L	0.5 %	Titration
	Indium standard solution	1000 mg/L	0.7 %	Titration
	Tellurium standard solution	1000 mg/L	1.5 %	Titration

Effective Date of Accreditation: 2023-10-21

	Vanadium standard solution	1000 mg/L	0.7 %	Titration
	Silver standard solution	1000 mg/L	0.4 %	Titration
	Beryllium standard solution	1000 mg/L	0.7 %	ICP-OES
	Titanium standard solution	1000 mg/L	0.4 %	ICP-OES
	Silicon standard solution	1000 mg/L	0.8 %	ICP-OES
	Yttrium standard solution	1000 mg/L	0.5 %	Titration
	Ammonium ion standard solution	1000 mg/L	0.8 %	Titration
	Bromide ion standard solution	1000 mg/L	0.6 %	Titration
	Chloride ion standard solution	1000 mg/L	0.6 %	Titration
	Fluoride ion standard solution	1000 mg/L	0.7 %	Titration
	Nitrate ion standard solution	1000 mg/L	0.8 %	IC
		4430 mg/L	0.9 %	
	Nitrite ion standard solution	1000 mg/L	0.8 %	IC
		3280 mg/L	0.9 %	
	Phosphate ion standard solution	1000 mg/L	0.7 %	IC
		3070 mg/L	0.7 %	IC
	Sulfate ion standard solution	1000 mg/L	0.9%	IC
		3000 mg/L	1.1 %	IC
	Chlorate ion standard solution	1000 mg/L	0.7 %	IC
	Bromate ion standard solution	2000 mg/L	0.5 %	IC
	Chlorite ion standard solution	1000 mg/L	1.9 %	IC
	Formaldehyde standard solution	1000 mg/L	4.8 %	GC
	Total organic carbon standard solution	1000 mg/L	0.6 %	HPLC
	Heptaoxyethylene dodecyl ether standard solution	100 mg/L	3.9 %	HPLC

Standard Solutions except pH Standard Solutions	5 Anion mixture standard solution			
	Fluoride ion standard solution	100 mg/L	0.5 %	IC
	Chloride ion standard solution	200 mg/L	1.3 %	
	Nitrite ion standard solution	33 mg/L	1.1 %	
	Chlorate ion standard solution	100 mg/L	0.8 %	
	Nitrate ion standard solution	443 mg/L	1.3 %	
	7 Anion mixture standard solution			
	Fluoride ion standard solution	20 mg/L	2.2 %	IC
	Chloride ion standard solution	20 mg/L	1.7 %	
	Nitrite ion standard solution	100 mg/L	1.1 %	
	Bromide ion standard solution	100 mg/L	2.3 %	
	Nitrate ion standard solution	100 mg/L	1.5 %	
	Phosphate ion standard solution	200 mg/L	2.1 %	
	Sulfate ion standard solution	100 mg/L	1.7 %	

Effective Date of Accreditation: 2023-10-21

Standard Solutions except pH Standard Solutions	25 VOC mixture standard solution			GC
	** Any combination of constituents listed as below can be supplied.			
	1, 1-Dichloroethylene	1000 mg/L	1.6 %	
	Dichloromethane	1000 mg/L	1.2 %	
	<i>trans</i> -1, 2- Dichloroethylene	1000 mg/L	1.1 %	
	<i>tert</i> -Butyl methyl ether	1000 mg/L	0.9 %	
	<i>cis</i> -1, 2-Dichloroethylene	1000 mg/L	1.1 %	
	Chloroform	1000 mg/L	0.9 %	
	1, 1, 1-Trichloroethane	1000 mg/L	1.1 %	
	Carbon tetrachloride	1000 mg/L	1.4 %	
	Benzene	1000 mg/L	1.2 %	
	1, 2-Dichloroethane	1000 mg/L	1.3 %	
	Trichloroethylene	1000 mg/L	1.1 %	
	1, 2-Dichloropropane	1000 mg/L	1.0 %	
	Bromodichloromethane	1000 mg/L	1.0 %	
	<i>cis</i> -1, 3- Dichloropropene	1000 mg/L	2.8 %	
	Toluene	1000 mg/L	1.0 %	
	<i>trans</i> -1, 3-Dichloropropene	1000 mg/L	4.1 %	
	Tetrachloroethylene	1000 mg/L	1.0 %	
	Dibromochloromethane	1000 mg/L	1.1 %	
	<i>o</i> -Xylene	1000 mg/L	1.1 %	
	Tribromomethane	1000 mg/L	1.1 %	
	1, 4-Dichlorobenzene	1000 mg/L	1.0 %	
	1, 4-dioxane	1000 mg/L	2.5 %	
	<i>p</i> -Xylene	1000 mg/L	0.7 %	
	<i>m</i> -Xylene	1000 mg/L	0.8 %	
	1, 1, 2-Trichloroethane	1000 mg/L	0.8 %	
	2 Musty odor substances mixture standard solution			
	2-Methylisoborneol	100 mg/L	2.2 %	GC
	Geosmin	100 mg/L	1.9 %	
	6 Phenols mixture standard solution			
	2-Chlorophenol	1000 mg/L	2.1 %	GC
	Phenol	1000 mg/L	2.2 %	
	2,6-Dichlorophenol	1000 mg/L	2.3 %	
	2,4-Dichlorophenol	1000 mg/L	2.3 %	
	2,4,6-Trichlorophenol	1000 mg/L	2.4 %	
	4-Chlorophenol	1000 mg/L	2.3 %	
	4 Haloacetic acid mixture standard solution			
	Chloroacetic acid	1000 mg/L	2.8 %	HPLC
	Dichloroacetic acid	1000 mg/L	3.0 %	
	Bromoacetic acid	1000 mg/L	2.8 %	
	Trichloroacetic acid	1000 mg/L	3.1 %	
	5 Anionic surfactants mixture standard solution			
	Sodium Decylbenzenesulfonate	100 mg/L	4.8 %	HPLC
	Sodium Undecylbenzenesulfonate	100 mg/L	4.9 %	
	Sodium Dodecylbenzenesulfonate	100 mg/L	4.9 %	
	Sodium Tridecylbenzenesulfonate	100 mg/L	5.2 %	
	Sodium Tetradecylbenzenesulfonate	100 mg/L	5.0 %	

*1 relative value

*2 IC: Ion chromatography

GC: Gas chromatography

HPLC: High-performance liquid chromatography

ICP-OES: Inductively coupled plasma optical emission spectrometry